

Product Name Guava ViaCount[®] Reagent
Part Number 4000-0040
Quantity 40 mL (100 Tests)

Product Description ViaCount Reagent contains two (2) fluorescent nucleic acid stains in a buffered, physiologically compatible saline solution. Sodium azide is included as a preservative at a concentration of 0.05%.

Lot Number 09-0161
Expiration Date 07 Oct 2009

	Specification	Result
Appearance	Faintly reddish-purple solution	Meets specification
pH	7.15 – 7.25	Meets specification

Test definition Performance testing on known mixtures of viable and non-viable Jurkat cells, against benchmark lot.

	Specification	Result
Fluorescence intensity	PM1 = Benchmark \pm 20%	Meets specification
Total Cell Count	PM2 = Benchmark \pm 20%	Meets specification
Live Cell Count	Benchmark \pm 10%	Meets specification
% Viability	Benchmark \pm 10%	Meets specification

Storage Conditions 2 to 8° C. Do not freeze.

Signature



Approval Signature:
Title : Director, Quality Systems
Date: 13 May 2009