
Certificate of Analysis

Total Lck Beadmates™

(100 Assay Points)

Catalog # 46-617

Lot # 24584

Components

Beadlyte® Anti-Lck Beads, Catalog # 42-617, Lot # 24584. One vial containing **125µl** of anti-Lck IgG conjugated to Luminex® Bead # 20 at **2,000 beads/µl (20X)** in a proprietary formulation of Tris-buffered salts and animal protein containing 0.05% sodium azide as a preservative.

Beadlyte® Anti-Lck, Biotin, Catalog # 44-617, Lot # 24584. One vial containing **125µl** of anti-Lck IgG (**20X**) in a proprietary formulation of Tris-buffered salts and animal protein containing 0.05% sodium azide as a preservative.

Specificity: Recognizes human and mouse Lck.

Applications: Optimal antibody pair for detection of total Lck. To be used in conjunction with Beadlyte® Cell Signaling Buffer Kits (Catalog #s 48-600 or 48-601).

Storage and Stability: Stable for 1 year at 4°C from date of shipment. Store in the **dark**.

FOR RESEARCH USE ONLY
NOT RECOMMENDED OR INTENDED FOR DIAGNOSIS OF DISEASE IN HUMANS OR ANIMALS
DO NOT USE IN HUMANS OR IN ANIMALS

Total Lck Beadmate™ Description

Use: The Beadlyte® Total Lck Beadmate™ pair is used in conjunction with Beadlyte® Cell Signaling Buffer Kits (Catalog #s 48-600 or 48-601) to detect the presence of Lck in cell lysates using the Luminex® xMAP™ system. Each Beadmate™ pair is ordered individually and can be combined for simultaneous multiplex analysis of cellular events. Beadlyte® Cell Signaling Buffer Kits are ordered separately and consist of a common set of reagents needed for using Beadmates™. The detection assay is a rapid, convenient alternative to Western Blotting and immunoprecipitation procedures. Each kit contains sufficient reagents for 100 individual assays.

Important note: For a detailed protocol on Cell Signaling Detection Procedures please see the COA (select the highest lot number) for the Beadlyte® Cell Signaling Buffer Kit available at:

<http://www.upstate.com/browse/productdetail.asp?ProductId=48-600>

Other components required but not included as part of kit are:

- Cell lysates or cell extracts harboring protein(s) of interest
- Vortex mixer
- Plate shaker
- Timer
- Variable volume (5-200 μ l) pipette + tips
- Sonication Bath (Catalog # 40-002)
- Millipore multiscreen vacuum manifold (Catalog # MAVM0960R)
- Luminex[®] 100[™] System
- Beadlyte[®] Cell Signaling Buffer Kit (Catalog # 48-600) or Cell Signaling Universal Buffer Kit (Catalog # 48-601)

Detection Protocol Summary

The assay procedure is a simple fluorescent bead-based sandwich immunoassay that is sensitive and easy to perform. A cell lysate or other sample is incubated overnight with beads coupled to a Lck specific antibody. The beads are washed and mixed with a biotinylated Lck specific reporter, followed by streptavidin-phycoerythrin. The amount of Lck is then quantified using the Luminex[®] xMAP[™] System.

Pre-wet filter plate and add 25 μ l of diluted cell lysate to each well with 25 μ l of 1X Lck bead solution.



*Overnight; dark
(4°C, shaking)*

Wash with 100 μ l Beadlyte[®] Cell Signaling Assay Buffer and add 25 μ l of 1X Lck reporter solution.




*1 hour; dark
(RT, shaking)*

Remove reporter and add 25 μ l diluted Beadlyte[®] Streptavidin-Phycoerythrin.




*30 min; dark
(RT, shaking)*


Remove Streptavidin-Phycoerythrin and resuspend in 100 μ l **Beadlyte[®] Cell Signaling Assay Buffer 1** and read results.




1. Bead with capture antibody



2. Capture antibody binds target proteins



3. Biotinylated reporter binds proteins



4. Streptavidin-PE binds biotinylated reporter antibody and emits fluorescent signal

Preparations for Assay Protocol

Single-plex analysis

The recommended lysis and assay buffers for a single-plex analysis of Total Lck Beadmates™ are Beadlyte® Cell Signaling **Lysis Buffer A or C** (Catalog #s 43-018 or 43-020) and Beadlyte® Cell Signaling **Assay Buffer 2** (Catalog # 43-011). Both buffers are included in the Beadlyte® Cell Signaling Buffer Kit (Catalog # 48-600). For the cell signaling assay and cell lysis protocols refer to the Beadlyte® Cell Signaling Buffer Kit COA (select the highest lot number) at: <http://www.upstate.com/browse/productdetail.asp?ProductId=48-600>.

Multiplex analysis

The recommended lysis and assay buffers multiplexing Total Lck Beadmates™ with other Beadmates™ are Beadlyte® Cell Signaling Universal Lysis Buffer (Catalog # 43-040) and Cell Signaling Universal Assay Buffer (Catalog # 43-041). Both buffers are included in the Beadlyte® Cell Signaling Universal Buffer Kit (Catalog # 48-601). For the cell signaling assay and cell lysis protocols refer to the Beadlyte® Cell Signaling Universal Buffer Kit COA (select the highest lot number) at: <http://www.upstate.com/browse/productdetail.asp?ProductId=48-601>.

For multiplexing Total Lck with other Beadmates™ that are *not* compatible with the Universal Buffer System, select the optimal lysis and assay buffers using the Buffer Selection Table in the Beadlyte® Cell Signaling Buffer Kit COA (Catalog # 48-600). The cell signaling assay and cell lysis protocols are also provided in the Beadlyte® Cell Signaling Buffer Kit COA at: <http://www.upstate.com/browse/productdetail.asp?ProductId=48-600> (select the highest lot number).

Note: Phospho and Total Beadmates should not be multiplexed together.

Representative Data:





Figure 1. Beadlyte[®] detection of Total Lck proteins in Jurkat T cell lysate. Jurkat T cells were grown to 90% confluence, serum starved overnight, and stimulated (at 80% confluence) with (●) or without (○) 5mM H₂O₂ for 15 minutes. Increasing amounts of cell lysate (lysed in Beadlyte[®] Cell Signaling Lysis Buffer C with protease inhibitors) were incubated overnight at 4°C with Beadlyte[®] Anti-Lck Beads. The Beads were washed and mixed at room temperature with Beadlyte[®] Anti-Total-Lck, Biotin, followed by streptavidin-PE. The Median Fluorescence Intensity (MFI) was measured using the Luminex[®] 100™ system.

Figure 2. Western blot detection of Total Lck in Jurkat T cell lysate. Jurkat T cells were grown, serum starved overnight, and stimulated with (H₂O₂) or without (NT) 5mM H₂O₂ for 15 minutes. Cells were lysed in Beadlyte[®] Cell Signaling Lysis Buffer C. 10µg of total lysate proteins per well were separated by SDS-PAGE, transferred to nitrocellulose, and probed with anti-Lck. Blots were incubated with an HRP-labeled secondary antibody, and visualized via chemiluminescence.



End-User License Agreement

By purchasing this product, which contains fluorescently labeled microsphere beads authorized by Luminex Corporation, you, the customer, acquire the right under Luminex Corporation's patent rights, if any, to use this product or any portion of this product, including without limitation the microsphere beads contained herein, only with Luminex's laser-based fluorescent analytical test instrumentation marketed under the name Luminex[®] 100™. One or more of the following US patents covers this product and the use thereof: #6,046,807, #5,981,180.