

**Certificate of Analysis**

28820 Single Oak Drive • Temecula, CA 92590  
Technical Support: T: 800 437-7500 • F: 800 437-7502  
www.millipore.com

**Acetyl-Histone H3 Immunoprecipitation (ChIP) Assay Kit**

Catalog # 17-245

Lot # DAM1520472

**Kit Components**

**Anti acetyl-Histone H3**, Catalog # 06-599 and 06-599-MN, see page two for more information. One vial containing **200 µg** of protein A purified IgG in **200 µl** and one vial containing **25 µg** of protein A purified IgG in **25 µL** in 0.1M Tris-glycine, 0.15M NaCl with 0.05% sodium azide.

**Salmon Sperm DNA/Protein A Agarose**, Catalog # 16-157C. One vial containing **600µg** sonicated Salmon Sperm DNA, 1.5mg BSA and approximately 4.5 mg recombinant Protein A. Provided as a 50% gel slurry for a final volume of **3 mL**. Suspended in TE Buffer containing 0.05% sodium azide.

**SDS Lysis Buffer**, Catalog # 20-163. One vial containing **10 ml** of 1% SDS, 10 mM EDTA, 50 mM Tris-HCl, pH 8.1.

**ChIP Dilution Buffer**, Catalog # 20-153. Two vials, each containing **24 mL** of 0.01% SDS, 1.1% Triton X-100, 1.2 mM EDTA, 16.7 mM Tris-HCl, pH 8.1, 167 mM NaCl.

**Low Salt Immune Complex Wash Buffer**, Catalog # 20-154. One vial containing **24 mL** of 0.1% SDS, 1% Triton X-100, 2 mM EDTA, 20 mM Tris-HCl, pH 8.1, 150 mM NaCl.

**High Salt Immune Complex Wash Buffer**, Catalog # 20-155. One vial containing **24 mL** of 0.1% SDS, 1% Triton X-100, 2 mM EDTA, 20 mM Tris-HCl, pH 8.1, 500 mM NaCl.

**LiCl Immune Complex Wash Buffer**, Catalog # 20-156. One vial containing **24 mL** of 0.25 M LiCl, 1% NP40, 1% deoxycholate, 1 mM EDTA, 10 mM Tris-HCl, pH 8.1.

**TE Buffer**, Catalog # 20-157. Two vials, each containing **24 mL** of 10 mM Tris-HCl, 1 mM EDTA, pH 8.0.

**5 M NaCl**, Catalog # 20-159. One vial containing **500 µL** of 5 M NaCl.

**0.5 M EDTA**, Catalog # 20-158. One vial containing **250 µL** of 0.5M EDTA, pH 8.0.

**1 M Tris-HCl, pH 6.5**, Catalog # 20-160. One vial containing **500 µL** of 1 M Tris-HCl, pH 6.5.

**FOR RESEARCH USE ONLY  
NOT FOR USE IN HUMANS**

**Kit Description**

**Quantity:** 22 chromatin immunoprecipitation (ChIP) assays. **Note:** One vial of Catalog # 06-599-MN was added to allow for 22 ChIP assays using the tested concentrations of antibody.

**Storage and Stability:** Stable for 1 year at 4°C from date of shipment.

**Use:** The assay kit contains reagents optimized for immunoprecipitation of transcriptionally active chromatin from mammalian cells using anti-acetyl-Histone H3. Detection of the gene or promoter of interest in immunoprecipitated chromatin must be empirically determined by the researcher. Quantitative PCR or Southern slot-blot analysis, using promoter specific primers or probe, is recommended.

Unless otherwise stated in our catalog or other company documentation accompanying the product(s), our products are intended for research use only and are not to be used for any other purpose, which includes but is not limited to, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses or any type of consumption or application to humans or animals.

### Technical Information for Kit Components

#### Anti-acetyl-Histone H3 (rabbit polyclonal IgG)

**Immunogen:** KLH-conjugated peptide corresponding to amino acids 1-20 of *Tetrahymena* histone H3 with a C-terminal cysteine added (ARTKQTAR<sub>[ACK]</sub>STG-G<sub>[ACK]</sub>APRKQL-C), where <sub>ACK</sub> denotes acetyl-lysine.

**Specificity:** Recognizes and is specific for acetylated histone H3.

**Species Cross-reactivity:** Human, mouse and *Tetrahymena*. Although other species not tested, a broad species cross reactivity is expected.

**Physical Form:** Frozen solution. Stable for one year at 4°C from date of shipment or 2 years at -20 °C Aliquot to avoid repeated freezing and thawing.

---


#### Application References:

1. Luo, R.X., *et al.*, *Cell*, 92: 463-473, 1998
2. Braunstein, M., *et al.*, *Mol. Cell. Biol.*, 16: 4349-4356, 1996.
3. Braunstein, M., *et al.*, *Genes & Devel.*, 7: 592-604, 1993.
4. Alberts, A. S., *et al.*, *Cell*, 92: 475-487, 1998

---

#### Quality Control Testing

**Chromatin Immunoprecipitation:** 10 µg of anti-acetyl Histone H3 immunoprecipitated acetylated histone H3 as determined by subsequent immunoblot analysis using 0.2 µg/mL antibody to detect histone from both 3T3/A31 and HeLa cells treated overnight with 5 mM sodium butyrate.



**Chromatin Immunoprecipitation:** Representative blot from a previous lot. Anti-acetyl-Histone H3 was used to immunoprecipitate chromatin from lysates of quiescent 3T3/A31 cells (lane 1) or sodium butyrate treated 3T3/A31 cells (lane 2). The immunoprecipitates were resolved by electrophoresis, transferred to nitrocellulose and probed with anti-acetyl-Histone H3 at 0.2 µg/mL. Proteins were visualized using a goat anti-rabbit secondary antibody conjugated to HRP and a chemiluminescence detection system. Arrow indicated acetylated Histone H3 (17 kDa).

Unless otherwise stated in our catalog or other company documentation accompanying the product(s), our products are intended for research use only and are not to be used for any other purpose, which includes but is not limited to, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses or any type of consumption or application to humans or animals.

Other components required but not included as part of kit are:

#### Reagents

- Cells, stimulated or treated as appropriate for the experimental system
- 37% Formaldehyde
- Protease inhibitors (PMSF, Aprotinin, Pepstatin A)
- PBS
- Dry ice
- Elution buffer: 1% SDS, 0.1M NaHCO<sub>3</sub>
- Molecular Biology grade Proteinase K
- Glycogen or tRNA
- 50% Phenol/50%Chloroform (containing 1% isoamyl alcohol)
- 95% Ethanol
- 70% Ethanol

#### Equipment

- vortex mixer
- rotating wheel/platform
- shaking incubator
- timer
- variable volume (5-1000µl) pipettes + tips
- microfuge
- variable temperature water bath
- humidified 37°C incubator
- cell scraper
- sonicator

---

**Safety Warnings and Precautions:** Formaldehyde should be used in a certified fume hood. Culture dishes containing formaldehyde should be placed in a sealed bag if returned to a humidified 37°C incubator to prevent exposing other cells in the incubator to formaldehyde. Ear protection must be worn during sonication. All chemicals should be considered potentially hazardous and handled with the principles of good laboratory practice.

---

## Chromatin Immunoprecipitation Protocol

### Part A. Optimization of DNA Shearing

Establish optimal conditions required for shearing cross-linked DNA to 200-1000 base pairs in length by following steps 1-9 below. Vary the power setting and/or the number of 10-second pulses during sonication of the samples. Be sure to keep the sample on ice at all times (the sonication generates heat which will denature the DNA). Check the size of sonicated DNA by gel electrophoresis after reversion of cross-links. Our experience shows DNA is sheared to the appropriate length with 3-4 sets of 10-second pulses using a Cole Parmer, High Intensity Ultrasonic Processor/Sonicator, 50-watt model equipped with a 2mm tip and set to 30% of maximum power. **Once sonication conditions have been optimized, keep cell number consistent for subsequent experiments. The protocol below for the optimization of DNA Shearing is for one Chip assay (~1 x 10<sup>6</sup> cells per condition).**

**Note:** Steps 3- 7 should be done on ice.

1. Stimulate or treat 1 x 10<sup>6</sup> cells on a 10cm dish as appropriate. (Cells should be treated under conditions for which transcriptional activation of the gene of interest has been demonstrated). Include one extra dish (1 X 10<sup>6</sup> cells) to be used solely for estimation of cell number.
2. Cross link histones to DNA by adding formaldehyde directly to culture medium to a final concentration of 1% and incubate for 10 minutes at 37°C. (For example, add 270µl 37% formaldehyde into 10 mL of growth medium on plate).
3. Aspirate medium, removing as much medium as possible. Wash cells twice using ice cold PBS containing protease inhibitors (1 mM phenylmethylsulfonyl fluoride (PMSF), 1 µg/mL aprotinin and 1µg/mL pepstatin A). **Note:** Add protease inhibitors to PBS just prior to use. PMSF has a half-life of approximately 30 minutes in aqueous solutions.
4. Scrape cells into conical tube.
5. Pellet cells for 4 minutes at 2000 x g at 4°C. Warm **SDS Lysis Buffer (Catalog # 20-163)** to room temperature to dissolve precipitated SDS and add protease inhibitors (inhibitors: 1 mM PMSF, 1 µg/mL aprotinin and 1 µg/mL pepstatin A).

Unless otherwise stated in our catalog or other company documentation accompanying the product(s), our products are intended for research use only and are not to be used for any other purpose, which includes but is not limited to, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses or any type of consumption or application to humans or animals.

6. Resuspend cell pellet in 200  $\mu$ L of **SDS Lysis Buffer (Catalog # 20-163)** and incubate for 10 minutes on ice. **Note:** The 200  $\mu$ L of SDS Lysis Buffer is per  $1 \times 10^6$  cells; if more cells are used, the resuspended cell pellet should be divided into 200  $\mu$ L aliquots so that each 200  $\mu$ L aliquot contains  $\sim 1 \times 10^6$  cells.
7. Sonicate lysate to shear DNA to lengths between 200 and 1000 basepairs being sure to keep samples ice cold (**Note:** Once sonication conditions have been optimized following steps 1 to 9, proceed to Part B, step 1 below).
8. Add 8  $\mu$ L **5 M NaCl (Catalog # 20-159)** and reverse crosslinks at 65°C for 4 hours.
9. Recover DNA by phenol/chloroform extraction and run sample (example 5  $\mu$ L, 10  $\mu$ L and 20  $\mu$ L samples) in an agarose gel to visualize shearing efficiency.

#### Part B. Experimental protocol.

*If sonication conditions have been optimized (Part A), complete steps 1 through 7 and continue with the protocol below. For a negative/background control, prepare a sample to use as a no-antibody immunoprecipitation control in step 5 below. Additionally, transcriptionally unactivated DNA samples should be prepared as controls for PCR in section II.*

1. Centrifuge samples (part A, step 7) for 10 minutes at 13,000 rpm at 4°C, and add 200  $\mu$ L of the sonicated cell supernatant to a new 2 mL-microcentrifuge tube. Discard pellet.
2. Dilute the sonicated cell supernatant 10 fold in **ChIP Dilution Buffer (Catalog # 20-153)**, adding protease inhibitors as above. This is done by adding 1800  $\mu$ L ChIP Dilution Buffer to the 200  $\mu$ L sonicated cell supernatant for a final volume of 2 mL in each immunoprecipitation condition. **Note:** *If proceeding to PCR a portion of the diluted cell supernatant 1% ( $\sim 20$   $\mu$ L) can be kept to quantitate the amount of DNA present in different samples at the PCR protocol, Part B, section II, step 6. This sample is considered to be your input/starting material and needs to have the Histone-DNA crosslinks reversed by heating at 65°C for 4 hours (see section II, step 3.).*
3. To reduce nonspecific background, pre-clear the 2 mL diluted cell supernatant with 80  $\mu$ L of **Salmon Sperm DNA/Protein A Agarose-50% Slurry (Catalog # 16-157C)** for 30 minutes at 4°C with agitation.
4. Pellet agarose by brief centrifugation and collect the supernatant fraction.
5. Add the immunoprecipitating antibody (the amount will vary per antibody) to the 2 mL supernatant fraction and incubate overnight at 4°C with rotation. *For a negative control, perform a no-antibody immunoprecipitation by incubating the supernatant fraction with 60  $\mu$ L of **Salmon Sperm DNA/Protein A Agarose-50% Slurry (Catalog # 16-157C)** for one hour at 4°C with rotation and proceed to step 7.*
6. Add 60  $\mu$ L of **Salmon Sperm DNA/Protein A Agarose Slurry (Catalog # 16-157C)** for one hour at 4°C with rotation to collect the antibody/histone complex.
7. Pellet agarose by gentle centrifugation (700 to 1000 rpm at 4°C,  $\sim 1$ min). Carefully remove the supernatant that contains unbound, non-specific DNA. Wash the protein A agarose/antibody/histone complex for 3-5 minutes on a rotating platform with 1ml of each of the buffers listed in the order as given below:
  - a) Low Salt Immune Complex Wash Buffer (Catalog # 20-154), **one wash**
  - b) High Salt Immune Complex Wash Buffer (Catalog # 20-155), **one wash**
  - c) LiCl Immune Complex Wash Buffer (Catalog # 20-156), **one wash**
  - d) TE Buffer (Catalog # 20-157), **two washes**

**After step 7 above, the sample is now a protein A/antibody/histone/DNA complex ready for either an Immunoprecipitation/Immunoblot assay (Section I) or Polymerase Chain Reaction (PCR) assay (Section II):**

#### Section I. Immunoprecipitation/Immunoblot protocol to detect histone.

- 1 Following washing of the beads in part B, step 7, immunoprecipitated histones can be analyzed by immunoblot analysis. Add 25  $\mu$ L of 1X Laemmli buffer per sample and boil for 10 minutes. Centrifuge briefly (3 seconds @ 10,000rpm) to pellet agarose beads. Load 20  $\mu$ L of supernatant per lane and perform immunoblot procedure as described per appropriate antibody.

Unless otherwise stated in our catalog or other company documentation accompanying the product(s), our products are intended for research use only and are not to be used for any other purpose, which includes but is not limited to, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses or any type of consumption or application to humans or animals.

**Section II. PCR protocol to amplify DNA that is bound to the immunoprecipitated histone.**

1. Freshly prepare elution buffer (1%SDS, 0.1M NaHCO<sub>3</sub>).
2. Elute the histone complex from the antibody by adding 250 µL elution buffer to the pelleted protein A agarose/antibody/histone complex from step 7d above. Vortex briefly to mix and incubate at room temperature for 15 minutes with rotation. Spin down agarose, and carefully transfer the supernatant fraction (eluate) to another tube and repeat elution. Combine eluates (total volume = ~500 µL).
3. Add 20 µL 5 M NaCl (Catalog # 20-159) to the combined eluates (500 µL) and reverse histone-DNA crosslinks by heating at 65°C for 4 hours. At this step the sample can be stored and -20°C and the protocol continued the next day.

**Note:** Include the input/starting material (*the sample saved from Part B, step 2, which has had the Histone-DNA crosslinks reversed*) as well as a transcriptionally-unactivated DNA sample as negative and background controls for the PCR reaction. *Previously, a 5 µL sample has been used in a nested PCR reaction. However, the amount of sample used per reaction must be determined empirically (e.g., titrate the sample at this step by using 1, 2, 5, or 10 µL per PCR reaction).* If PCR results are poor, complete steps 4, 5 and 6 below to purify the DNA sample. NOTE: Handle the samples carefully, some DNA may be lost during the purification steps.

4. Add 10 µL of 0.5 M EDTA (Catalog # 20-158), 20 µL 1 M Tris-HCl, pH 6.5 (Catalog # 20-160) and 2 µL of 10 mg/mL Proteinase K to the combined eluates and incubate for one hour at 45°C.
5. Recover DNA by phenol/chloroform extraction and ethanol precipitation. Addition of an inert carrier, such as 20 µg glycogen or yeast tRNA, helps visualize the DNA pellet. Wash pellets with 70% ethanol and air dry.
6. Resuspend pellets in an appropriate buffer for PCR or slot-blot reactions. PCR or slot-blot conditions must be determined empirically.

Unless otherwise stated in our catalog or other company documentation accompanying the product(s), our products are intended for research use only and are not to be used for any other purpose, which includes but is not limited to, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses or any type of consumption or application to humans or animals.