

Product Name Guava ViaCount[®] Reagent
Part Number 4000-0040
Quantity 40 mL (100 Tests)

Product Description ViaCount Reagent contains two (2) fluorescent nucleic acid stains in a buffered, physiologically compatible saline solution. Sodium azide is included as a preservative at a concentration of 0.05%.

Lot Number 09-0448
Expiration Date 15 Sep 2010

| | Specification | Result |
|-------------------|---------------------------------|---------------------|
| Appearance | Faintly reddish-purple solution | Meets specification |
| pH | 7.15 – 7.25 | Meets specification |

Test definition Performance testing on known mixtures of viable and non-viable Jurkat cells, against benchmark lot.

| | Specification | Result |
|-------------------------------|---------------------------|---------------------|
| Fluorescence intensity | PM1 = Benchmark \pm 20% | Meets specification |
| Total Cell Count | PM2 = Benchmark \pm 20% | Meets specification |
| Live Cell Count | Benchmark \pm 10% | Meets specification |
| % Viability | Benchmark \pm 10% | Meets specification |

Storage Conditions 2 to 8° C. Do not freeze.

Signature



Approval Signature:

Title: Site Quality Manager

Date: 7 October 2009