

Aervent® Cartridge Filter

0.2 µm Rated

Catalogue Number: CTGR75S01

Lot Number: C9CN77953

Good Manufacturing Practices

This product was manufactured in a Millipore facility which adheres to Good Manufacturing Practices.

ISO® 9001 Quality Standard

This product was manufactured in a Millipore facility whose Quality Management System is approved by an accredited registering body to the appropriate ISO 9001 Quality Systems Standard.

Non-Fiber Releasing

This product was manufactured with a PTFE membrane which meets the criteria for a "non-fiber releasing" filter as defined in 21 CFR 210.3 (b) (6).

Component Materials Toxicity

Component materials were tested and meet the criteria for the USP Class VI Biological Test for Plastics.

100% Integrity Testing in Manufacturing

Each unit must pass the Millipore Integrity Test correlated to the *Brevundimonas diminuta* ASTM® F838 bacterial challenge test.

Quality Assurance Lot Release Criteria

This manufacturing lot was sampled, tested and released by Quality Assurance to the following specifications:

Bacterial Retention

Samples were quantitatively retentive of a minimum *Brevundimonas diminuta* challenge concentration of 1×10^7 CFU/cm² using ASTM F838 methodology.

USP Bacterial Endotoxins

A cartridge aqueous extraction contains less than 0.5 EU/mL as determined using the Limulus Amebocyte lysate (LAL) test.

Integrity

Samples exhibited a bubble point equal to or greater than 1.6 psig (1.103 mbar) in a 70/30% IPA/water mixture using nitrogen as the test gas at 23° C.

Samples exhibited a nitrogen diffusional flow rate of less than or equal to 1.2 cc/min at 14 psig (966 mbar) in a 70/30% IPA/water mixture at 23° C, per 5-inch cartridge.

Samples exhibited a HydroCarr test value less than or equal to 0.38 mL/min per 5-inch cartridge at 38 psig (2620 mbar).

Thermal and Hydraulic Stress

Samples were steamed in place at 145° C and maintained integrity after a forward stress to 80 psid (5.5 bar) and a reverse stress to 60 psid (4.1 bar).

Air Flow Rate and Pressure Drop

Samples met a maximum pressure drop of 2.8 psid (190 mbar) at an air flow rate of 50 SCFM (8.5 nm³/hr), and 0 psig 0 (mbar) outlet air pressure at 23° C per 5-inch cartridge.

Quality Assurance Audit Criteria

This product was designed and manufactured to meet the following specifications. Performance is confirmed by testing on an audit basis.

Toxicity

This product is non-toxic per the current USP General (Mouse) Safety Test.

Gravimetric Extractables

The extractables level was equal to or less than 18 mg per 5-inch cartridge after 24 hours in a 70/30% IPA/water mixture at controlled room temperature.

Multiple Sterilization Cycles

Integrity was maintained after 100 forward and 50 reverse steam-in-place cycles of 30 minutes at 145° C.



Peter Eichert
Quality Manager

Millipore and Aervent are registered trademarks of Millipore Corporation. ASTM is a registered trademark of the American Society for Testing and Materials. ISO is a registered trademark of The International Organization for Standardization. P3.5320 Rev F 07/07

MILLIPORE