

**OptiScale<sup>®</sup> Capsule Filter  
with Hydrophilic Durapore<sup>®</sup>  
Membrane**

Durapore 0.1 µm Rated  
Catalogue Number: SVVLA47FF3  
Lot Number: C9NN55557

**Good Manufacturing Practices**

This product was manufactured in a Millipore facility which adheres to Good Manufacturing Practices.

**ISO<sup>®</sup> 9001 Quality Standard**

This product was manufactured in a Millipore facility whose Quality Management System is approved by an accredited registering body to the appropriate ISO 9001 Quality Systems Standard.

**Indirect Food Additive**

All component materials meet the FDA Indirect Food Additive requirements cited in 21 CFR 177-182.

**Quality Assurance Lot  
Release Criteria**

This manufacturing lot was sampled, tested and released by Quality Assurance to the following specifications:

**Integrity**

Samples exhibited a water bubble point equal to or greater than 70.0 psig (4830 mbar) with air at 23 °C.

**Quality Performance Criteria**

This product was designed and manufactured to meet the following specifications:

**Oxidizable Substances**

Affluent meets the requirements for USP Sterile Water for Injection after a water flush of 100 mL per autoclaved capsule.

**Toxicity**

Component materials were tested and meet the criteria for the USP <88> Biological Reactivity Tests for Class VI Plastics.

This product is non-toxic per the USP <88> Safety Test with 0.9% Sodium Chloride extraction procedure.

**Multiple Sterilization Cycles**

Capsule integrity was maintained after 3 autoclave cycles of 60 minutes at 126 °C.



Peter Eichert  
Quality Manager