

**OptiScale™ Capsule Filter with  
Prefiltration and Clarification  
Media**

Polysep II 1.0/0.5 µm Nominal Rated  
Catalogue Number: SGW6A47HH3  
Lot Number: C9CN81272

**Good Manufacturing Practices**

This product was manufactured in a Millipore facility which adheres to Good Manufacturing Practices.

**ISO 9000 Quality Standard**

This product was manufactured in a Millipore facility whose Quality Management System is approved by an accredited registering body to the appropriate ISO 9000 Quality Systems Standard.

**Indirect Food Additive**

All component materials meet the FDA Indirect Food Additive requirements cited in 21 CFR 177.182.

**Quality Performance Criteria**

This product was designed and manufactured to meet the following specifications.

**Oxidizable Substances**

Effluent meets the requirements for USP Sterile Water for Injection after a water flush of 100 mL per autoclaved capsule.

**Toxicity**

Component materials were tested and meet the criteria for the USP <88> Biological Reactivity Tests for Class VI Plastics.

This product is non-toxic per the USP <88> Safety Test with 0.9% Sodium Chloride extraction procedure.

**Multiple Sterilization Cycles**

Capsule integrity was maintained after 3 autoclave cycles of 60 minutes at 121 °C.



Peter Eichert  
BioPharmaceutical Quality Manager

MILLIPORE

Millipore is a registered trademark of Millipore Corporation.  
OptiScale is a trademark of Millipore Corporation.  
P60211 Rev B 08/08