

## Opticap® XL Capsule Filter with Milligard® Media

0.5/0.2 µm Nominal Rated  
Catalogue Number: KWSSA10HH1  
Lot Number: C9JN19872

### Good Manufacturing Practices

This product was manufactured in a Millipore facility which adheres to Good Manufacturing Practices.

### ISO® 9001 Quality Standard

This product was manufactured in a Millipore facility whose Quality Management System is approved by an accredited registering body to the appropriate ISO 9001 Quality Systems Standard.

### Non-Fiber Releasing

This product was manufactured with a Millipore media combination which meets the criteria for a "non-fiber releasing" filter as defined in 21 CFR 210.3 (b)(6).

### Component Materials Toxicity

Component materials were tested and meet the criteria for the USP Class VI Biological Test for Plastics.

### Indirect Food Additive

All component materials meet the FDA Indirect Food Additive requirements cited in 21 CFR 177-182.

### European Pressure Equipment Directive

Millipore Corporation certifies that this product complies with the European Pressure Equipment Directive, 97/23/EC of 29 May 1997. This product has been classified under Article 3 § 3 of the Pressure Vessel Directive. It has been designed and manufactured in accordance with sound engineering practice to ensure safe use. In compliance with Article 3 § 3 of this Pressure Equipment Directive, this product does not bear the CE mark.

### Quality Assurance Lot Release Criteria

This manufacturing lot was sampled, tested, and released by Quality Assurance to the following specifications:

#### USP Bacterial Endotoxins

A sample aqueous extraction contains less than 0.5 EU/ml as determined using the Limulus Amebocyte Lysate (LAL) test.

#### Structural Integrity

Samples from Milligard cartridges used in this lot were found to be integral as measured by an aerosol challenge test.

#### Flow Rate and Pressure Drop

Samples met a maximum pressure drop of 30.2 psid (2.08 bar) at 10.0 gpm (37.8 L/min) with clean water at 23° C.

#### USP Oxidizable Substances

Effluent was negative after a water flush of 5 l per autoclaved sample.

### Quality Performance Criteria

This product was designed and manufactured to meet the following specifications:

#### Toxicity

This product is non-toxic per the current USP General (Mouse) Safety Test.

#### Gravimetric Extractables

The extractables level was equal to or less than 55 mg per sample after 24 hours in ASTM® Type I reagent-grade water at controlled room temperature.

#### Multiple Sterilization Cycles

Sample integrity was maintained after 3 autoclave cycles of 30 minutes at 121° C.



Peter Eicherl  
Quality Manager



Millipore, Milligard and Opticap are registered trademarks of Millipore Corporation. ISO is a registered trademark of the International Organization for Standardization. ASIM is a registered trademark of the American Society for Testing and Materials.

P72039 Rev D 03/09