

## **OptiScale® Capsule Filter with Prefiltration and Clarification**

### **Media**

Polysep II 2.0/1.2 µm Nominal Rated

Catalogue Number: SGW9A47FH3

Lot Number: C9KN32353

### **Good Manufacturing Practices**

This product was manufactured in a Millipore facility which adheres to Good Manufacturing Practices.

### **ISO® 9001 Quality Standard**

This product was manufactured in a Millipore facility whose Quality Management System is approved by an accredited registering body to the appropriate ISO 9001 Quality Systems Standard.

### **Indirect Food Additive**

All component materials meet the FDA Indirect Food Additive requirements cited in 21 CFR 177-182.

### **Quality Performance Criteria**

This product was designed and manufactured to meet the following specifications.

### **Oxidizable Substances**

Effluent meets the requirements for USP Sterile Water for Injection after a water flush of 100 mL per autoclaved capsule.

### **Toxicity**

Component materials were tested and meet the criteria for the USP <88> Biological Reactivity Tests for Class VI Plastics.

This product is non-toxic per the USP <88> Safety Test with 0.9% Sodium Chloride extraction procedure.

### **Multiple Sterilization Cycles**

Capsule integrity was maintained after 3 autoclave cycles of 60 minutes at 121 °C.



Peter Eichert  
Quality Manager

Millipore and OptiScale are registered trademarks of Millipore Corporation.  
ISO is a registered trademark of The International Organization for Standardization  
P60211 Rev C 06/09