

**OptiScale® Capsule Filter with
Hydrophilic Millipore®
Express® SHC Membrane**

Millipore Express SHC 0.5/0.2 µm Rated
Catalogue Number: SHGEA47FH3
Lot Number: C9JN18873

Good Manufacturing Practices

This product was manufactured in a Millipore facility which adheres to Good Manufacturing Practices.

ISO® 9001 Quality Standard

This product was manufactured in a Millipore facility whose Quality Management System is approved by an accredited registering body to the appropriate ISO 9001 Quality Systems Standard.

Indirect Food Additive

All component materials meet the FDA Indirect Food Additive requirements cited in 21 CFR 177-182.

Millipore, Millipore Express and OptiScale are registered trademarks of Millipore Corporation. ISO is a registered trademark of the International Standards Organization. P88060 Rev C 06/09

Quality Performance Criteria

This product was designed and manufactured to meet the following specifications:

Oxidizable Substances

Effluent meets the requirements for USP Sterile Water for Injection after a water-fush of 100 mL per autoclaved capsule.

Toxicity

Component materials were tested and meet the criteria for the USP <88> Biological Reactivity Tests for Class V Plastics.

This product is non-toxic per the USP <88> Safety Test with 0.9% Sodium Chloride extraction procedure.

Multiple Sterilization Cycles

Capsule integrity was maintained after 3 autoclave cycles of 60 minutes at 126 °C.



Peter Eichart
Quality Manager