

# Durapore® Hydrophilic Cartridge Filter

0.45 µm Rated

Catalogue Number: CVHL02TP3

Lot Number: C9KN33908

## Good Manufacturing Practices

This product was manufactured in a Millipore facility which meets or exceeds FDA Device Good Manufacturing Practice standards.

## ISO® 9001 Quality Standard

This product was manufactured in a Millipore facility whose Quality Management System is approved by an accredited registering body to the appropriate ISO 9001 Quality Systems Standard.

## Non-Fiber Releasing

This product was manufactured with a Millipore Durapore membrane which meets the criteria for a "non-fiber releasing" filter as defined in 21 CFR 210.3 (b) (6).

## Component Materials Toxicity

Component materials were tested and meet the criteria for the USP Class VI Biological Test for Plastics.

## Indirect Food Additive

The Durapore membrane used in this product meets the FDA Indirect Food Additive requirements cited in 21 CFR 177.2910. All other component materials also meet the FDA Indirect Food Additive requirements cited in 21 CFR 177-182.

## 100% Integrity Testing in Manufacturing

Each unit must pass the Millipore Integrity Test.

Millipore and Durapore are registered trademarks of Millipore Corporation. ISO is a registered trademark of The International Organization for Standardization. ASTM is a registered trademark of the American Society for Testing and Materials.

Rev D 05/09

## Quality Assurance Lot Release Criteria

This manufacturing lot was sampled, tested and released by Quality Assurance to the following specifications:

### USP Bacterial Endotoxins

A cartridge aqueous extraction contains less than 0.5 EU/ml as determined using the Limulus Amebocyte lysate (LAL) test.

### Integrity

Samples exhibited a water bubble point equal to or greater than 28 psig (1930 mbar) with air at 23 °C. Samples exhibited an air diffusional flow rate of less than or equal to 30 cc/min at 22.0 psig (1520 mbar) in water at 23 °C per 20-inch cartridge.

### Thermal and Hydraulic Stress

Samples were steamed in place at 121 °C and maintained integrity after a forward stress to 80 psid (5.5 bar) and a reverse stress to 50 psid (3.4 bar).

### Flow Rate and Pressure Drop

Samples met a maximum pressure drop of 2.0 psid (140 mbar) at 4.0 gpm (15.1 L/min) per 20-inch cartridge with clean water at 23 °C.

### USP Oxidizable Substances

Effluent was negative after a water flush of 2000 mL per autoclaved 20-inch cartridge.

## Quality Assurance Audit Criteria

This product was designed and manufactured to meet the following specifications. Performance is confirmed by testing on an audit basis.

### Toxicity

This product is non-toxic per the current USP General (Mouse) Safety Test.

### Gravimetric Extractables

The extractables level was equal to or less than 45 mg per 10-inch cartridge after 24 hours in ASTM® Type 1 reagent-grade water at controlled room temperature.

### Multiple Sterilization Cycles

Integrity was maintained after 10 steam-in-place cycles of 30 minutes at 121 °C.



Peter Eichert  
Jaffrey Quality Manager



Brigitte Weber  
Molsheim Quality Manager

