

## **OptiScale® Capsule Filter with Hydrophilic Durapore® Membrane**

Durapore 0.1 µm Rated  
Catalogue Number: SVLA47FH3  
Lot Number: C9MNA43060

### **Good Manufacturing Practices**

This product was manufactured in a Millipore facility which adheres to Good Manufacturing Practices.

### **ISO® 9001 Quality Standard**

This product was manufactured in a Millipore facility whose Quality Management System is approved by an accredited registering body to the appropriate ISO 9001 Quality Systems Standard.

### **Indirect Food Additive**

All component materials meet the FDA Indirect Food Additive requirements cited in 21 CFR 177.182.

## **Quality Assurance Lot Release Criteria**

This manufacturing lot was sampled, tested and released by Quality Assurance to the following specifications:

### **Integrity**

Samples exhibited a water bubble point equal to or greater than 70.0 psig (4830 mbar) with air at 23 °C.

## **Quality Performance Criteria**

This product was designed and manufactured to meet the following specifications:

### **Oxidizable Substances**

Affluent meets the requirements for USP Sterile Water for injection after a water flush of 100 mL per autoclaved capsule.

### **Toxicity**

Component materials were tested and meet the criteria for the USP <88> Biological Reactivity Tests for Class V Plastics.

This product is non-toxic per the USP <88> Safety Test with 0.9% Sodium Chloride extraction procedure.

### **Multiple Sterilization Cycles**

Capsule integrity was maintained after 3 autoclave cycles of 60 minutes at 126 °C.



Peter Eichert  
Quality Manager

Millipore, Durapore and OptiScale are registered trademarks of Millipore Corporation. ISO is a registered trademark of The International Organization for Standardization P60210 Rev C 06/09