



Certificate of Quality

Item Description: KTGR01TC1 1/PK
Millipore Catalog Number: SN2M027A6
Millipore Lot Number: C9JN17313

Millipore Corporation certifies that the above-referenced product:

- Is a Custom Order and although it is not a standard Millipore catalogue item it has been manufactured with the same due care .
- Was 100% integrity tested
- Was sampled, tested and released by Quality Assurance to specifications including :
 - USP Bacterial Endotoxins: < 0.5 EU/mL as determined by the LAL test
 - Bubble Point: ≥ 16 psi in a 70/30% IPA / water mixture when tested with nitrogen
 - Nitrogen Diffusion Rate: ≤ 22 cc/min at 14 psig (wet dynamically with 70/30% IPA/water)
 - Housing Burst: ≥ 200 psig
 - Air Pressure Drop: ≤ 1.05 psid at 50 SCFM air flow rate and 0 psig outlet air pressure. (This correlates to a 1.2 psid pressure drop at 100 SCFM at 30 psig inlet.)

Cindy Kadlec
Quality Assurance

Date Approved: 12-Oct-2009

The format of this document has been updated to include approval date, page numbers and/or print date.

THIS DOCUMENT IS PROPRIETARY AND ITS CONTENTS ARE THE EXCLUSIVE PROPERTY OF MILLIPORE CORPORATION.
THIS DOCUMENT MAY NOT BE ALTERED IN ANY FORM WHATSOEVER, WITHOUT PRIOR WRITTEN PERMISSION FROM MILLIPORE CORPORATION'S QUALITY SYSTEMS MANAGER.